

Schedule

OPUS Precision Instruments Pte Ltd
Blk 512 Chai Chee Lane
#05-13
Bedok Industrial Estate
Singapore 469028

Certificate No. : LA-2005-0324-C
Issue No. : 23
Date : 07 July 2025
Expiry of Certificate : 10 May 2026
Page : 1 of 24

FIELD OF TESTING: Calibration and Measurement

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
A. DIMENSIONAL METROLOGY		
A1. Bevel Protractor (Lab) i Straightness & Parallelism ii Indication of Error	BS 1685 : 2008 Opus in-house procedure OPCP-01 Issue 04	0.0001 inch 5 min
A2. Bore Gauge (Lab) 0 ~ 2 mm (Plunger Travel)	JIS B 7515 : 1982 Opus in-house procedure OPCP-02 Issue 04	2.0 µm
A3. Vernier Caliper (Lab) 0 ~ 1000 mm	JIS B 7507 : 2016 Opus in-house procedure OPCP-03 Issue 04	10 µm
A4. Caliper Checker (Lab) Up to 600 mm Parallelism	Opus in-house procedure OPCP-04 Issue 04	1.3 µm 1.2 µm
A5. Dial Caliper Gauge / Dial Thickness Gauge (Lab) Up to 50 mm (Digital) Up to 50 mm (Analog)	Opus in-house procedure OPCP-05 Issue 04	1.1 µm 2.2 µm
A6. Dial Gauge Tester (Lab) Up to 50.8 mm Up to 2 inch	Opus in-house procedure OPCP-06 Issue 04	0.3 µm 0.00001 inch
A7. Dial Indicator (Lab) up to 10 mm above 10 mm to 20 mm above 20 mm to 50 mm above 50 mm to 80 mm above 80 mm to 100 mm	JIS B 7503: 2017 DIN 879-1: 1999 ASME B89.1.10M-2001 Opus in-house procedure OPCP-07 Issue 04	0.5 µm 1.0 µm 1.4 µm 2.0 µm 2.5 µm

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 2 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
A8. Dial Test Indicator (Lab) 0 ~ 1 mm	JIS B 7533: 2015 ASME B89.1.10M-2001 Opus in-house procedure OPCP-08 Issue 04	0.6 µm
A9. Digimatic Indicator (Lab) Up to 30 mm Above 30 mm to 50 mm Above 50 mm to 80 mm Above 80 mm to 100 mm	Opus in-house procedure OPCP-09 Issue 04	0.8 µm 1.4 µm 2.0 µm 2.5 µm
A10. Feeler Gauge (Lab) 0 ~ 1 mm	BS 957: 2008 Opus in-house procedure OPCP-10 Issue 04	1.0 µm
A11. Granite Surface Plate (Lab / on-site) Up to 3000 mm i Overall Flatness ii Variation from local flatness	BS 817: 2008 Opus in-house procedure OPCP-12 Issue 04	3.0 µm 1.0 µm
A12. Linear Height Gauge (Lab / on-site) 0 to 600 mm 600 to 900 mm	Opus in-house procedure OPCP-13 Issue 04	2.0 µm 3.0 µm
A13. Vernier Height Gauge (Lab) up to 600 mm	Opus in-house procedure OPCP-14 Issue 04	10 µm
A14. Height Setting Micrometer (Lab) up to 300mm	ISO 7863: 1984(R2018) Opus in-house procedure OPCP-15 Issue 04	1.3 µm
A15. Holtest (Lab) 3 mm to 63 mm 63 mm to 100 mm 100 mm to 200 mm	DIN 863 Part 4: 1999 & Opus in-house procedure OPCP-16 Issue 04	2.3 µm 2.4 µm 3.8 µm

Schedule



Certificate No. : LA-2005-0324-C Issue No. : 23

Date : 07 July 2025 Page : 3 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
A16. External Micrometer (Lab) Up to 100 mm Above 100 mm to 300 mm Above 300 mm to 500 mm Above 500 mm to 1000mm	Opus in-house procedure OPCP-17 Issue 04	1.0 µm 2.0 µm 3.0 µm 4.0 µm
A17. Internal & Stick Micrometer (Lab) i. Micrometer Head – up to 25mm ii. Extension Rod- Up to 300mm	BS 959: 2008 Opus in-house procedure OPCP-18 Issue 04	2.0 µm 4.0 µm
A18. Depth Micrometer (Lab) Up to 300mm	BS 6468: 2008 Opus in-house procedure OPCP-19 Issue 04	2.0 µm
A19. Pin Gauge / Plug Gauge (Lab) Up to 25 mm Up to 50 mm Up to 100 mm	BS 969: 2008, as a guide Opus in-house procedure OPCP-20 Issue 04	0.5 µm 1.0 µm 1.2 µm
A20. Plain Ring Gauge (Lab) 3 mm to 30 mm Above 30 mm to 100 mm Above 100mm to 150mm	BS 969: 2008, as a guide Opus in-house procedure OPCP-21 Issue 04	0.8 µm 1.2 µm 2.6 um
A21. Profile Projector (Lab / on-site) Up to 300 mm	JIS B 7184: 1999 Opus in-house procedure OPCP-22 Issue 04	3.0 µm
A22. Setting Rod for External Micrometer (Lab) Up to 500 mm Above 500 mm to 1000 mm	Opus in-house procedure OPCP-24 Issue 04	1.2 µm 2.0 µm
A23. Thread Plug Gauge (Lab) Up to 50 mm i Pitch & Major Diameter ii Pitch iii Flank Angle	ISO 1502: 1996 ANSI / ASME B1.2: 1983(R2017) BS 919 Pt 1 to Pt 4: 2007 BS 3643 Pt 1 & 2: 2007 BS 1580 Pt 1 & 3: 2007 Opus in-house procedure OPCP-26 Issue 04	1.0 µm 3.0 µm 3.0 min

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 4 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
A24. Toolmaker Microscope (Lab / on-site) Up to 300 mm	JIS B 7153: 1995 Opus in-house procedure OPCP-27 Issue 04	3.0 μ m
A25. Universal Length Measuring (Lab / on-site) 0 to 10 mm (10 to 50) mm (50 to 80) mm (80 to 100) mm (100 to 500) mm (500 to 1000) mm	Opus in-house procedure OPCP-31 Issue 04	0.2 μ m 0.3 μ m 0.4 μ m 0.5 μ m 2.0 μ m 4.0 μ m
A26. Vernier Depth Gauge (Lab) 0 ~ 600mm	BS 6365: 2008 Opus in-house procedure OPCP-32 Issue04	10 μ m
A27. Coating Thickness Foils (Plastic) (Lab)	Opus In-house procedure OPCP36 Issue 04	0.6 μ m
A28. CMM (Coordinate Measuring Machine) (Lab) (400x710x420) mm	ISO 10360-2 2009 AS GUIDE Opus in-house procedure OPCP-25 Issue 04	4.8 μ m
A29. Jig & Fixture Measurements (Lab) (400x710x420) mm	Opus in-house procedure OPCP-66 Issue 04	8.0 μ m
A30. Coating Thickness Gauge (Lab) 0 ~ 2900 μ m	Opus in-house procedure OPCP-35 Issue 04	0.7 μ m

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 5 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
<p>B. FORCE MEASUREMENT</p> <p>B1. Push / Pull Gauge (Lab) up to 500 gf above 0.5 kgf to 5 kgf above 5 kgf to 10 kgf above 10 kgf to 50 kgf</p> <p>B2. Torque Gauge / Torque Meter (Lab) up to 1.5 kgf-cm up to 3.6 kgf-cm up to 9 kgf-cm up to 15 kgf-cm</p> <p>B3. Weighing Scale (Lab/ On-site) up to 310 g, Resolution: 0.0001g up to 620 g, Resolution: 0.001g up to 5.2 kg, Resolution: 0.01g up to 30 kg, Resolution: 1g up to 100 kg, Resolution: 10g up to 150 kg, Resolution: 20g</p> <p>B4. Hand Torque Tool (Lab) Type I 0.5 to 10 N.m 10 to 50 N.m 50 to 100 N.m 100 to 1000 N.m 1000 to 1500 N.m Type II 0.5 to 10 N.m 10 to 50 N.m 50 to 100 N.m 100 to 1000 N.m 1000 to 1500 N.m</p> <p>B5. Velocity - Optical Tachometer Measuring (Non-Contact Tachometer) 240 ~ 1000 RPM 1000 ~ 5000 RPM 5000 ~ 10000 RPM 10000 ~ 30000 RPM 30000 ~ 60000 RPM</p>	<p>Opus in-house procedure OPCP-23 Issue 04</p> <p>Opus in-house procedure OPCP-28 Issue 04</p> <p>Opus in-house procedure OPCP-30 Issue 04</p> <p>ISO 6789 -1& 2 : 2017 Opus in-house procedure OPCP-29 Issue 04</p> <p>Opus in-house procedure OPCP-059 Issue 04</p>	<p>0.2 gf 0.002 kgf 0.01 kgf 0.02 kgf</p> <p>0.004 kgf-cm 0.01 kgf-cm 0.02 kgf-cm 0.1 kgf-cm</p> <p>0.0018 g 0.003 g 0.05 g 1.0 g 10 g 20 g</p> <p>1.1 % reading 0.2 % reading 0.3 % reading 0.4 % reading 0.3 % reading</p> <p>1.5 % reading 0.4 % reading 0.5 % reading 0.7 % reading 0.3 % reading</p> <p>(Lab & Onsite) 2.0 RPM 2.0 RPM 2.0 RPM 2.0 RPM 3.0 RPM</p>

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 6 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
C. Pressure Instruments (Lab & on-Site)		
C1. Pressure Measuring Devices Pneumatic Pressure Gauges Chart/Pen recorders Digital Indicators Manometers, Data loggers, Compound gauge Oil free & Oxygen Gauges Oil Pressure Gauges -13 psi to 30 psi 0 to 300 psi 0 to 1000 psi 0 to 3000 psi 0 to 10000 psi	Opus in-house procedure OPCP-033 Issue 04	0.07 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S
C2. Pressure Switches (Pneumatic Service/ Oil free & Oxygen Service/ Hydraulic Oil) 0 to 30 psi 0 to 300 psi 0 to 1000 psi 0 to 3000 psi 0 to 10000 psi	Opus in-house procedure OPCP-034 Issue 04	0.07 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S
C3. Pressure Transmitter (Pneumatic Service/ Oil free & Oxygen Service/ Hydraulic Oil) 0 to 30 psi 0 to 300 psi 0 to 1000 psi 0 to 3000 psi 0 to 10000 psi	Opus in-house procedure OPCP-037 Issue 04	0.12 % of F.S 0.12 % of F.S 0.12 % of F.S 0.12 % of F.S 0.12 % of F.S
C4. Pressure Transducer (Pneumatic Service/ Oil free & Oxygen Service/ Hydraulic Oil) 0 to 30 psi 0 to 300 psi 0 to 1000 psi 0 to 3000 psi 0 to 10000 psi	Opus in-house procedure OPCP-037 Issue 04	0.04 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S 0.03 % of F.S

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 7 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
D. Electrical (Lab & on-Site)		
D1. DC VOLTAGE MEASURING	Opus in-house procedure OPCP-053 Issue 04	
0 ~ 50 mV		0.006 mV
(50 ~ 100) mV		0.0077 mV
(100 ~ 150) mV		0.01 mV
(150 ~ 202) mV		0.012 mV
(0.20 ~ 0.25) V		0.000015 V
(0.25 ~ 0.50) V		0.000024 V
(0.50 ~ 1.00) V		0.000041 V
(1.00 ~ 2.02) V		0.00008 V
(2.00 ~ 10.00) V		0.00034 V
(10.00 ~ 20.20) V		0.0007 V
(20 ~ 100) V		0.004 V
(100 ~ 202) V		0.008 V
(200 ~ 500) V		0.023 V
(500 ~ 1000) V		0.04 V
D2. DC CURRENT MEASURING	Opus in-house procedure OPCP-049 Issue 04	
0 ~ 50 μ A		0.041 μ A
(50 ~ 100) μ A		0.047 μ A
(100 ~ 150) μ A		0.053 μ A
(150 ~ 202) μ A		0.059 μ A
(0.200 ~ 0.250) mA		0.00007 mA
(0.250 ~ 0.500) mA		0.000093 mA
(0.500 ~ 1.000) mA		0.00014 mA
(1.00 ~ 2.02) mA		0.00024 mA
(2.00 ~ 10.00) mA		0.00093 mA
(10.00 ~ 20.2) mA		0.0017 mA
(20 ~ 100) mA		0.013 mA
(100 ~ 202) mA		0.024 mA
(0.200 ~ 1.000) A		0.00022 A
(1.000 ~ 2.020) A		0.00041 A
(2 ~ 10) A		0.0051 A
(10 ~ 20) A		0.011 A
(20 ~ 30) A		0.1 A

Schedule



Certificate No. : LA-2005-0324-C Issue No. : 23

Date : 07 July 2025 Page : 8 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
D3. AC VOLTAGE MEASURING	Opus in-house procedure OPCP-045 Issue 04	
(20 ~ 50) mV	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz	0.19 mV 0.063 mV 0.23 mV 2.6 mV
(50 ~ 100) mV	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz	0.3 mV 0.082 mV 0.26 mV 2.6 mV
(100 ~ 150) mV	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz	0.41 mV 0.11 mV 0.29 mV 2.6 mV
(150 ~ 202) mV	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz	0.53 mV 0.13 mV 0.33 mV 2.7 mV
(0.200 ~ 0.250) V	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz 100 to 500 kHz	0.00099 V 0.00046 V 0.0023 V 0.0038 V 0.026 V
(0.250 ~ 0.500) V	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz 100 to 500 kHz	0.0016 V 0.00053 V 0.0023 V 0.0044 V 0.026 V
(0.500 ~ 1.000) V	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz 100 to 500 kHz	0.0028 V 0.0007 V 0.0025 V 0.0057 V 0.027 V
(1.00 ~ 2.020) V	10 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz 20 to 99.999 kHz 100 to 500 kHz	0.0052 V 0.0012 V 0.0032 V 0.0085 V 0.03 V
(2.00 ~ 10.00) V	40 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz	0.027 V 0.0086 V 0.022 V

Schedule



Certificate No. : LA-2005-0324-C Issue No. : 23

Date : 07 July 2025 Page : 9 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
(10.0 ~ 20.2) V	40 to 44 Hz 45 to 999 Hz 1 to 19.999 kHz	0.051 V 0.012 V 0.027 V
(20.0 ~ 100) V	40 to 44 Hz 45 to 999 Hz	0.096 V 0.13 V
(100 ~ 202) V	40 to 44 Hz 45 to 999 Hz	0.2 V 0.16 V
(200 ~ 500) V	40 to 44 Hz 45 to 999 Hz	0.71 V 0.48 V
(500 ~ 1000) V	46 to 999 Hz 1 to 10 kHz	0.76 V 2.3 V
D4. AC CURRENT MEASURING	Opus in-house procedure OPCP-041 Issue 04	
(20 ~ 50) μ A	40 to 44 Hz 45 to 999 Hz	0.49 μ A 0.34 μ A
(50 ~ 100) μ A	40 to 44 Hz 45 to 999 Hz	0.59 μ A 0.38 μ A
(100 ~ 150) μ A	40 to 44 Hz 45 to 999 Hz	0.69 μ A 0.42 μ A
(150 ~ 202) μ A	40 to 44 Hz 45 to 999 Hz	0.8 μ A 0.46 μ A
(0.200 ~ 0.250) mA	40 to 44 Hz 45 to 999 Hz	0.0029 mA 0.00089 mA
(0.250 ~ 0.500) mA	40 to 44 Hz 45 to 999 Hz	0.0032 mA 0.0011 mA
(0.500 ~ 1.000) mA	40 to 44 Hz 45 to 999 Hz	0.0039 mA 0.0014 mA
(1.00 ~ 2.02) mA	40 to 44 Hz 45 to 999 Hz	0.0059 mA 0.002 mA

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 10 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
AC CURRENT MEASURING (Continue)	Opus in-house procedure OPCP-041 Issue 04	
(2 ~ 10) mA	40 to 44 Hz 45 to 999 Hz	0.029 mA 0.014 mA
(10 ~ 20.2) mA	40 to 44 Hz 45 to 999 Hz	0.059 mA 0.02 mA
(20 ~ 100) mA	40 to 44 Hz 45 to 999 Hz	0.39 mA 0.13 mA
(100 ~ 202) mA	40 to 44 Hz 45 to 999 Hz	0.59 mA 0.2 mA
(0.200 ~ 1.000) A	40 to 44 Hz 45 to 999 Hz	0.0039 A 0.0023 A
(1.000 ~ 2.02) A	40 to 44 Hz 45 to 999 Hz	0.0059 A 0.0031 A
(2.000 ~ 10) A	30 to 44 Hz 45 to 99 Hz 0.10 to 1 kHz	0.039 A 0.016 A 0.041 A
(10 ~ 20) A	30 to 44 Hz 45 to 99 Hz 0.10 to 1 kHz	0.053 A 0.025 A 0.075 A
(20 ~ 30) A	30 to 44 Hz 45 to 99 Hz 0.10 to 1 kHz	0.077 A 0.036 A 0.11 A

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 11 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
<p>D5. RESISTANCE MEASURING (passive) - 2 wire</p> <p>185 mΩ 280 mΩ 1.2502 Ω 10.191 Ω 100.212 Ω 1.000229 kΩ</p> <p>10.0005 kΩ 100.001 kΩ 0.999827 MΩ 9.99886 MΩ 99.943 MΩ 993.54 MΩ</p>	<p>Opus in-house procedure OPCP-038 Issue 04</p>	<p>47 mΩ 47 mΩ 0.047 Ω 0.048 Ω 0.052 Ω 0.00014 kΩ</p> <p>0.00097 kΩ 0.0094 kΩ 0.00017 MΩ 0.0046 MΩ 0.59 MΩ 12 MΩ</p>
<p>D6. RESISTANCE MEASURING (passive) - 4 Wire</p> <p>0 mΩ 100.00 mΩ 1.00367 Ω 10.0079 Ω 100.00365 Ω 1.000015 kΩ 10.000270 kΩ 100.0005 kΩ</p>	<p>Opus in-house procedure OPCP-039 Issue 04</p>	<p>5.8 mΩ 5.8 mΩ 0.0059 Ω 0.007 Ω 0.012 Ω 0.000093 kΩ 0.00093 kΩ 0.0093 kΩ</p>
<p>D7. CAPACITANCE MEASURING</p> <p>1.0014 nF 10.096 nF 20.031 nF 50.312 nF 99.71 nF 1.0001 μF 10.227 μF</p>	<p>Opus in-house procedure OPCP-046 Issue 04</p> <p>1 kHz</p>	<p>0.027 nF 0.06 nF 0.2 nF 0.3 nF 0.4 nF 0.005 μF 0.08 μF</p>

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 12 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
<p>D8. SIMULATED RESISTANCE MEASURING</p> <p>(10 ~ 100) Ω (0.100 ~ 1.000) kΩ (1 ~ 10) kΩ (10 ~ 100) kΩ (0.10 ~ 1) MΩ (1 ~ 10) MΩ</p>	<p>Opus in-house procedure OPCP-038 Issue 04</p>	<p>0.093 Ω 0.00041 kΩ 0.0036 kΩ 0.035 kΩ 0.00035 MΩ 0.0036 MΩ</p>
<p>D9. FREQUENCY MEASURING</p> <p>(3 ~ 10) Hz 10 Hz ~ 100 Hz 100 Hz ~ 1000 Hz 1 kHz ~ 10 kHz 10 kHz ~ 100 kHz 100 kHz ~ 1000 kHz</p>	<p>Opus in-house procedure OPCP-055 Issue 04</p>	<p>0.00024 Hz 0.0024 Hz 0.024 Hz 0.00024 kHz 0.0024 kHz 0.024 kHz</p>
<p>D10. DC CURRENT CLAMP MEASURING</p> <p>(0.1 ~ 0.5) A 0.5 A ~ 1 A (1 ~ 10) A (10 ~ 50) A (50 ~ 100) A (100 ~ 500) A (500 ~ 1000) A (1000 ~ 1500) A</p>	<p>Opus in-house procedure OPCP-049 Issue 04</p>	<p>0.12 A 0.12 A 0.17 A 0.4 A 0.9 A 4.0 A 6.0 A 9.0 A</p>
<p>D11. AC CURRENT CLAMP MEASURING</p> <p>(0.1 ~ 0.5) A (0.5 ~ 1) A (1 ~ 10) A (10 ~ 50) A (50 ~ 100) A (100 ~ 500) A (500 ~ 1000) A (1000 ~ 1500) A</p>	<p>Opus in-house procedure OPCP-041 Issue 04</p> <p>(30Hz to 60 Hz)</p>	<p>0.11 A 0.12 A 0.17 A 0.4 A 0.9 A 4.0 A 6.0 A 9.0 A</p>

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 13 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
D12. INSULATION RESISTANCE MEASURING (10 ~ 50) k Ω (50 ~ 100) k Ω (100 ~ 500) k Ω (0.50 ~ 1.00) M Ω (1 ~ 5) M Ω (5 ~ 10) M Ω (10 ~ 50) M Ω (50 ~ 100) M Ω (100 ~ 500) M Ω (500 ~ 1000) M Ω (1000 ~ 2000) M Ω	Opus in-house procedure OPCP-057 Issue 04	0.12 k Ω 0.24 k Ω 1.2 k Ω 0.0024 M Ω 0.012 M Ω 0.35 M Ω 1.8 M Ω 3.5 M Ω 18 M Ω 36 M Ω 70 M Ω
D13. INSULATION TEST VOLTAGE MEASURING 0 ~ 50 V 50 V ~ 100 V 100 V ~ 250 V 250 V ~ 500 V 500 V ~ 1000 V	Opus in-house procedure OPCP-058 Issue 04	0.32 V 0.61 V 1.5 V 3 V 5.8 V
D14. CONTINUITY RESISTANCE MEASURING 1 Ω 10 Ω 19 Ω 100 Ω 190 Ω 1000 Ω	Opus in-house procedure OPCP-047 Issue 04	0.061 Ω 0.081 Ω 0.11 Ω 0.29 Ω 0.5 Ω 2.4 Ω

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 14 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
D15. DC VOLTAGE SOURCING 0 ~ 100 mV (0.100 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 500) V (500 ~ 1000) V	Opus in-house procedure OPCP-052 Issue 04	0.0085 mV 0.000038 V 0.00034 V 0.0052 V 0.036 V 0.06 V
D16. AC VOLTAGE SOURCING (1 ~ 100) mV (0.100 ~ 1) V (1 ~ 10) V (10 ~ 100) V (100 ~ 500) V (500 ~ 1000) V	Opus in-house procedure OPCP-044 Issue 04 (10 Hz to 20 kHz)	0.17 mV 0.0011 V 0.011 V 0.11 V 0.64 V 1.1 V
D17. DC CURRENT SOURCING 0 ~ 100 µA (0.100 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 400) mA (0.400 ~ 1) A (1 ~ 3) A (3 ~ 10) A	Opus in-house procedure OPCP-048 Issue 04	0.088 µA 0.00064 mA 0.0082 mA 0.064 mA 0.26 mA 0.00082 A 0.0042 A 0.02 A
D18. AC CURRENT SOURCING (20 ~ 100) µA (0.100 ~ 1) mA (1 ~ 10) mA (10 ~ 100) mA (100 ~ 400) mA (0.400 ~ 1) A (1 ~ 3) A (3 ~ 10) A	Opus in-house procedure OPCP-040 Issue 04 (10 Hz to 1 kHz)	0.25 µA 0.0017 mA 0.025 mA 0.17 mA 0.94 mA 0.0017 A 0.0088 A 0.025 A

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 15 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
D19. RESISTANCE SOURCING (1 ~ 10) Ω (10 ~ 100) Ω (0.100 ~ 1) k Ω (1 ~ 10) k Ω (10 ~ 100) k Ω (0.1 ~ 1) M Ω (1 ~ 10) M Ω	Opus in-house procedure OPCP-061 Issue 04	0.0047 Ω 0.017 Ω 0.00013 k Ω 0.0013 k Ω 0.013 k Ω 0.00013 M Ω 0.0048 M Ω
D20. FREQUENCY SOURCING (3 ~ 1000 Hz) 100 Hz ~ 1000 Hz 1 kHz ~ 10 kHz (10 ~ 100) kHz 100 kHz ~ 1000 kHz	Opus in-house procedure OPCP-054 Issue 04	0.012 Hz 0.12 Hz 0.0012 kHz 0.012 kHz 0.12 kHz
D21. DC POWER MEASURING (0.3 ~ 10) W (10 ~ 100) W (100 ~ 500) W (0.500 ~ 1.000) kW 1 kW ~ 12 kW	Opus in-house procedure OPCP-051 Issue 04	0.011 W 0.2 W 0.99 W 0.0013 kW 0.01 kW
D22. AC POWER MEASURING (Power Factor = 1) (0.3 ~ 100 W) (10 ~ 100) W (100 ~ 500) W (0.500 ~ 1.000) kW 1 kW ~ 12 kW	Opus in-house procedure OPCP-043 Issue 04 (50 Hz & 400 Hz)	0.12 W 0.6 W 1.2 W 0.0018 kW 0.014 kW
D23. DC HIGH VOLTAGE SOURCING 0 ~ 1 kV (1 ~ 5) kV (5 ~ 10) kV (10 ~ 20) kV (20 ~ 30) kV (30 ~ 40) kV	Opus in-house procedure OPCP-050 Issue 04	0.09 kV 0.15 kV 0.24 kV 0.7 kV 1.5 kV 1.6 kV

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 16 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
<p>D24. AC HIGH VOLTAGE SOURCING (50 Hz)</p> <p>0.01 kV ~ 1 kV 1 kV ~ 5 kV 5 kV ~ 10 kV 10 kV ~ 20 kV 20 kV ~ 28 kV</p>	Opus in-house procedure OPCP-042 Issue 04	0.12 kV 0.32 kV 0.73 kV 1.5 kV 1.9 kV
<p>D25. Milli-Ohm meter / Continuity Tester</p> <p>10 mΩ 22 mΩ 30 mΩ 51 mΩ 100 mΩ 500 mΩ 1 Ω 2 Ω 5 Ω 10 Ω 20 Ω 50 Ω 100 Ω 200 Ω 500 Ω 1000 Ω</p>	Opus in-house procedure OPCP-056 Issue 04	0.86 mΩ 0.87 mΩ 0.99 mΩ 1.6 mΩ 2.6 mΩ 6.3 mΩ 0.071 Ω 0.079 Ω 0.096 Ω 0.13 Ω 0.26 Ω 0.58 Ω 1.2 Ω 3.3 Ω 5.8 Ω 12 Ω
<p>D26. High Voltage Insulation Resistance Measuring</p> <p>10 kΩ 500 kΩ 1 MΩ 2 MΩ 4 MΩ 5 MΩ 10 MΩ 20 MΩ 30 MΩ 40 MΩ 50 MΩ 100 MΩ 200 MΩ</p>	Opus in-house procedure OPCP-056 Issue 04	0.47 kΩ 5.8 kΩ 0.012 MΩ 0.024 MΩ 0.047 MΩ 0.068 MΩ 0.12 MΩ 0.24 MΩ 0.36 MΩ 0.5 MΩ 0.59 MΩ 1.4 MΩ 3.8 MΩ

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 17 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)													
500 MΩ 1000 MΩ 2000 MΩ 5000 MΩ		8.2 MΩ 14 MΩ 29 MΩ 69 MΩ													
E. Temperature															
E1. (a) Temperature Indicator/controller, Digital Thermometer & Temperature Measuring Instruments (electrical simulation)	Opus in-house procedure OPCP-062 Issue 04														
Thermocouple Simulation -Type J -210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C		<table border="1"> <thead> <tr> <th data-bbox="1200 936 1359 965">Lab</th> <th data-bbox="1359 936 1513 965">Onsite</th> </tr> </thead> <tbody> <tr> <td data-bbox="1200 965 1359 994">0.38 °C</td> <td data-bbox="1359 965 1513 994">1.2 °C</td> </tr> <tr> <td data-bbox="1200 994 1359 1023">0.30 °C</td> <td data-bbox="1359 994 1513 1023">1.2 °C</td> </tr> <tr> <td data-bbox="1200 1023 1359 1052">0.29 °C</td> <td data-bbox="1359 1023 1513 1052">1.0 °C</td> </tr> <tr> <td data-bbox="1200 1052 1359 1081">0.33 °C</td> <td data-bbox="1359 1052 1513 1081">1.0 °C</td> </tr> <tr> <td data-bbox="1200 1081 1359 1111">0.36 °C</td> <td data-bbox="1359 1081 1513 1111">1.1 °C</td> </tr> </tbody> </table>		Lab	Onsite	0.38 °C	1.2 °C	0.30 °C	1.2 °C	0.29 °C	1.0 °C	0.33 °C	1.0 °C	0.36 °C	1.1 °C
Lab	Onsite														
0.38 °C	1.2 °C														
0.30 °C	1.2 °C														
0.29 °C	1.0 °C														
0.33 °C	1.0 °C														
0.36 °C	1.1 °C														
Thermocouple Simulation -Type K -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1370 °C		<table border="1"> <tbody> <tr> <td data-bbox="1200 1173 1359 1202">0.42 °C</td> <td data-bbox="1359 1173 1513 1202">1.2 °C</td> </tr> <tr> <td data-bbox="1200 1202 1359 1232">0.32 °C</td> <td data-bbox="1359 1202 1513 1232">1.2 °C</td> </tr> <tr> <td data-bbox="1200 1232 1359 1261">0.30 °C</td> <td data-bbox="1359 1232 1513 1261">0.96 °C</td> </tr> <tr> <td data-bbox="1200 1261 1359 1290">0.35 °C</td> <td data-bbox="1359 1261 1513 1290">0.96 °C</td> </tr> <tr> <td data-bbox="1200 1290 1359 1319">0.40 °C</td> <td data-bbox="1359 1290 1513 1319">1.1 °C</td> </tr> </tbody> </table>		0.42 °C	1.2 °C	0.32 °C	1.2 °C	0.30 °C	0.96 °C	0.35 °C	0.96 °C	0.40 °C	1.1 °C		
0.42 °C	1.2 °C														
0.32 °C	1.2 °C														
0.30 °C	0.96 °C														
0.35 °C	0.96 °C														
0.40 °C	1.1 °C														
Thermocouple Simulation -Type T -250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C		<table border="1"> <tbody> <tr> <td data-bbox="1200 1411 1359 1440">0.76 °C</td> <td data-bbox="1359 1411 1513 1440">1.3 °C</td> </tr> <tr> <td data-bbox="1200 1440 1359 1469">0.29 °C</td> <td data-bbox="1359 1440 1513 1469">1.2 °C</td> </tr> <tr> <td data-bbox="1200 1469 1359 1498">0.29 °C</td> <td data-bbox="1359 1469 1513 1498">0.89 °C</td> </tr> <tr> <td data-bbox="1200 1498 1359 1527">0.3 °C</td> <td data-bbox="1359 1498 1513 1527">0.89 °C</td> </tr> </tbody> </table>		0.76 °C	1.3 °C	0.29 °C	1.2 °C	0.29 °C	0.89 °C	0.3 °C	0.89 °C				
0.76 °C	1.3 °C														
0.29 °C	1.2 °C														
0.29 °C	0.89 °C														
0.3 °C	0.89 °C														
Thermocouple Simulation -Type R -0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1760 °C		<table border="1"> <tbody> <tr> <td data-bbox="1200 1619 1359 1648">0.97 °C</td> <td data-bbox="1359 1619 1513 1648">2.2 °C</td> </tr> <tr> <td data-bbox="1200 1648 1359 1677">0.60 °C</td> <td data-bbox="1359 1648 1513 1677">2.0 °C</td> </tr> <tr> <td data-bbox="1200 1677 1359 1706">0.66 °C</td> <td data-bbox="1359 1677 1513 1706">2.0 °C</td> </tr> </tbody> </table>		0.97 °C	2.2 °C	0.60 °C	2.0 °C	0.66 °C	2.0 °C						
0.97 °C	2.2 °C														
0.60 °C	2.0 °C														
0.66 °C	2.0 °C														
Thermocouple Simulation -Type S 0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1760 °C		<table border="1"> <tbody> <tr> <td data-bbox="1200 1776 1359 1805">0.97 °C</td> <td data-bbox="1359 1776 1513 1805">2.2 °C</td> </tr> <tr> <td data-bbox="1200 1805 1359 1834">0.58 °C</td> <td data-bbox="1359 1805 1513 1834">2.0 °C</td> </tr> <tr> <td data-bbox="1200 1834 1359 1863">0.66 °C</td> <td data-bbox="1359 1834 1513 1863">2.0 °C</td> </tr> </tbody> </table>		0.97 °C	2.2 °C	0.58 °C	2.0 °C	0.66 °C	2.0 °C						
0.97 °C	2.2 °C														
0.58 °C	2.0 °C														
0.66 °C	2.0 °C														

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 18 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)	
Thermocouple Simulation -Type N -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C		Lab 0.56°C 0.36 °C 0.33 °C 0.32 °C 0.39 °C	Onsite 1.2 °C 0.92 °C 0.92 °C 0.92 °C 0.92 °C
Thermocouple Simulation - Type E -250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C		0.71 °C 0.29°C 0.29 °C 0.31 °C 0.32 °C	1.6 °C 1.5 °C 1.2 °C 1.2 °C 1.2 °C
(b) Temperature Calibrator / Temperature Sourcing Instruments (Electrical Simulation)	Opus in-house procedure OPCP-062 Issue 04		
Thermocouple Measuring - Type J -210 °C to -100 °C -100 °C to -30 °C -30 °C to 150 °C 150 °C to 760 °C 760 °C to 1200 °C		Lab 0.62 °C 0.58 °C 0.57 °C 0.59 °C 0.60 °C	Onsite 0.93 °C 0.93 °C 0.75 °C 0.82 °C 0.85 °C
Thermocouple Measuring - Type K -200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 1000 °C 1000 °C to 1370 °C		0.56 °C 0.49 °C 0.48 °C 0.51 °C 0.55 °C	1.1 °C 1.0 °C 0.78 °C 0.78 °C 1.0 °C
Thermocouple Measuring - Type T -250 °C to -150 °C -150 °C to 0 °C 0 °C to 120 °C 120 °C to 400 °C		0.91 °C 0.58 °C 0.58 °C 0.58 °C	1.1 °C 1.1 °C 0.73 °C 0.73 °C

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 19 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)	
		Lab	Onsite
<p>Thermocouple Measuring - Type R</p> <p>-0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1760 °C</p>		<p>1.1 °C 0.8 °C 0.9 °C</p>	<p>2.0 °C 1.8 °C 1.9 °C</p>
<p>Thermocouple Measuring - Type S</p> <p>0 °C to 250 °C 250 °C to 1000 °C 1000 °C to 1760 °C</p>		<p>1.3 °C 1.0 °C 1.1 °C</p>	<p>2.0 °C 1.8 °C 1.9 °C</p>
<p>Thermocouple Measuring - Type N</p> <p>-200 °C to -100 °C -100 °C to -25 °C -25 °C to 120 °C 120 °C to 410 °C 410 °C to 1300 °C</p>		<p>0.81 °C 0.68 °C 0.67 °C 0.67 °C 0.70 °C</p>	<p>1.4 °C 1.4 °C 0.96 °C 1.0 °C 1.0 °C</p>
<p>Thermocouple Measuring - Type E</p> <p>-250 °C to -100 °C -100 °C to -25 °C -25 °C to 350 °C 350 °C to 650 °C 650 °C to 1000 °C</p>		<p>0.87 °C 0.58 °C 0.57 °C 0.58 °C 0.59 °C</p>	<p>1.3 °C 0.73 °C 0.73 °C 0.73 °C 0.73 °C</p>

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 20 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)	
<p>E2. (a) RTD Indicator / Digital Thermometer</p> <p>(PRT100- 4 wire) -99.983 °C 0.021 °C 30.016 °C 60.020 °C 99.983 °C 199.96 °C 399.86 °C 799.36 °C</p> <p>(2 wire) PRT100 -200 °C to 0 °C 0 °C to 300 °C 300 °C to 800 °C</p> <p>PRT25 -200 °C to 0 °C 0 °C to 800 °C</p> <p>PRT250 -200 °C to 0 °C 0 °C to 800 °C</p> <p>PRT500 -200 °C to 0 °C 0 °C to 800 °C</p> <p>PRT1000 -200 °C to 0 °C 0 °C to 800 °C</p>	<p>Opus in-house procedure OPCP-063 Issue 04</p>	<p>Lab 0.06 °C 0.03 °C 0.04 °C 0.05 °C 0.06 °C 0.08 °C 0.09 °C 0.10 °C</p> <p>Lab Onsite 0.16 °C 0.24 °C 0.64 °C 0.42 °C 0.64 °C 0.61 °C</p> <p>Lab 0.58 °C 0.70 °C</p> <p>Lab 0.29 °C 0.35 °C</p> <p>Lab 0.12 °C 1.1 °C</p> <p>Lab 0.10 °C 0.52 °C</p>	
<p>(b) RTD Calibrator/Sourcing instruments (Simulation)</p> <p>(PRT100-4 wire) -200 °C to 0 °C 0 °C to 200 °C 200 °C to 600 °C 600 °C to 800 °C</p> <p>(PRT100-2 wire) -200 °C to 0 °C 0 °C to 200 °C 200 °C to 600 °C 600 °C to 800 °C</p>	<p>Opus in-house procedure OPCP-063 Issue 04</p>	<p>Lab Onsite 0.11 °C 0.27 °C 0.16 °C 0.31 °C 0.27 °C 0.42 °C 0.27 °C 0.53 °C</p> <p>Lab Onsite 0.17 °C 0.26 °C 0.18 °C 0.71 °C 0.22 °C 0.77 °C 0.23 °C 0.84 °C</p>	

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 21 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)																																				
<p>E3. (a) Contact thermometer with/ without display (RTD Probe/ RTD probe c/w display)</p> <p>-30 °C to 0 °C 0 °C to 125 °C 125 °C to 250 °C 250 °C to 400 °C</p> <p>(b) Bimetal thermometer/ Temperature transmitter/ Recorder/ Switch</p> <p>-30 °C to 0 °C 0 °C to 125 °C 125 °C to 250 °C 250 °C to 400 °C</p>	<p>Opus in-house procedure OPCP-064 Issue 04</p>	<table border="0"> <tr> <td>Lab</td> <td>Onsite</td> </tr> <tr> <td>0.16 °C</td> <td>0.16 °C</td> </tr> <tr> <td>0.16 °C</td> <td>0.16 °C</td> </tr> <tr> <td>0.25 °C</td> <td>0.25 °C</td> </tr> <tr> <td>0.64 °C</td> <td>0.64 °C</td> </tr> <tr> <td>Lab</td> <td>Onsite</td> </tr> <tr> <td>0.18 °C</td> <td>0.18 °C</td> </tr> <tr> <td>0.25 °C</td> <td>0.25 °C</td> </tr> <tr> <td>0.38 °C</td> <td>0.38 °C</td> </tr> <tr> <td>0.70 °C</td> <td>0.70 °C</td> </tr> </table>	Lab	Onsite	0.16 °C	0.16 °C	0.16 °C	0.16 °C	0.25 °C	0.25 °C	0.64 °C	0.64 °C	Lab	Onsite	0.18 °C	0.18 °C	0.25 °C	0.25 °C	0.38 °C	0.38 °C	0.70 °C	0.70 °C																
Lab	Onsite																																					
0.16 °C	0.16 °C																																					
0.16 °C	0.16 °C																																					
0.25 °C	0.25 °C																																					
0.64 °C	0.64 °C																																					
Lab	Onsite																																					
0.18 °C	0.18 °C																																					
0.25 °C	0.25 °C																																					
0.38 °C	0.38 °C																																					
0.70 °C	0.70 °C																																					
<p>E4. (a) Thermocouple with/ without display</p> <p>-30 °C to 0 °C 0 °C to 125 °C 125 °C to 250 °C 250 °C to 400 °C</p> <p>Thermocouple with display</p> <p>-30 °C to 0 °C 0 °C to 125 °C 125 °C to 250 °C 250 °C to 400 °C</p> <p>Thermocouple without display</p> <p>K Type</p> <p>-30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p> <p>J Type</p> <p>-30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p>	<p>Opus in-house procedure OPCP-065 Issue 04</p> <p>Opus in-house procedure OPCP-065 Issue 04</p> <p>Opus in-house procedure OPCP-065 Issue 04</p>	<table border="0"> <tr> <td>Lab</td> <td></td> </tr> <tr> <td>0.31 °C</td> <td></td> </tr> <tr> <td>0.34 °C</td> <td></td> </tr> <tr> <td>0.40 °C</td> <td></td> </tr> <tr> <td>0.78 °C</td> <td></td> </tr> <tr> <td></td> <td>Onsite</td> </tr> <tr> <td></td> <td>0.22 °C</td> </tr> <tr> <td></td> <td>0.30 °C</td> </tr> <tr> <td></td> <td>0.35 °C</td> </tr> <tr> <td></td> <td>0.73 °C</td> </tr> <tr> <td></td> <td>Onsite</td> </tr> <tr> <td></td> <td>0.95 °C</td> </tr> <tr> <td></td> <td>0.57 °C</td> </tr> <tr> <td></td> <td>0.86 °C</td> </tr> <tr> <td></td> <td>Onsite</td> </tr> <tr> <td></td> <td>0.72 °C</td> </tr> <tr> <td></td> <td>0.57 °C</td> </tr> <tr> <td></td> <td>0.86 °C</td> </tr> </table>	Lab		0.31 °C		0.34 °C		0.40 °C		0.78 °C			Onsite		0.22 °C		0.30 °C		0.35 °C		0.73 °C		Onsite		0.95 °C		0.57 °C		0.86 °C		Onsite		0.72 °C		0.57 °C		0.86 °C
Lab																																						
0.31 °C																																						
0.34 °C																																						
0.40 °C																																						
0.78 °C																																						
	Onsite																																					
	0.22 °C																																					
	0.30 °C																																					
	0.35 °C																																					
	0.73 °C																																					
	Onsite																																					
	0.95 °C																																					
	0.57 °C																																					
	0.86 °C																																					
	Onsite																																					
	0.72 °C																																					
	0.57 °C																																					
	0.86 °C																																					

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 22 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)	
<p>T Type -30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p> <p>R/ S Type -30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p> <p>N Type -30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p> <p>E Type -30 °C to 0 °C 0 °C to 250 °C 250 °C to 400 °C</p>		<p>Onsite 0.95 °C 0.57 °C 0.86 °C</p> <p>Onsite 1.6 °C 1.7 °C 1.8 °C</p> <p>Onsite 1.2 °C 0.77 °C 1.0 °C</p> <p>Onsite 0.51 °C 0.57 °C 0.86 °C</p>	
<p>(b) Thermocouple extension wire (Thermocouple simulation wire)</p> <p>K TYPE 0 °C to 50 °C J TYPE 0 °C to 50 °C T TYPE 0 °C to 50 °C R TYPE 0 °C to 50 °C S TYPE 0 °C to 50 °C N TYPE 0 °C to 50 °C E TYPE 0 °C to 50 °C</p>	Opus in-house procedure OPCP-065 Issue 04	<p>Lab</p> <p>6.4 µV 7.3 µV 6.5 µV 4.8 µV 4.8 µV 5.6 µV 8.2 µV</p>	
<p>E5. Dry block calibrator/ Liquid bath</p> <p>-30 °C to 125 °C 125 °C to 250 °C 250 °C to 500 °C</p>	Opus in-house procedure OPCP-067 Issue 04	<p>Lab</p> <p>0.15 °C 0.20 °C 0.61 °C</p>	<p>Onsite</p> <p>0.15 °C 0.20 °C 0.61 °C</p>
<p>E6. Temperature & Humidity Measurement (Thermohygrometer/Barometer/ Humidity transmitter/Recorder/ Thermohygrograph)</p> <p>10 °C to 50 °C 30 %r.h to 90% r.h.</p>	Opus in-house procedure OPCP-068 Issue 04	<p>Lab</p> <p>0.5 °C 3.0 % relative humidity</p>	

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 23 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)	
E7. (a) Surface Probe with/without display Up to 50 °C 50 °C to 150 °C 150 °C to 200 °C 200 °C to 350 °C (b) Hot Plate Up to 50 °C 50 °C to 150 °C 150 °C to 350 °C	Opus in-house procedure OPCP-069 Issue 04 Opus in-house procedure OPCP-069 Issue 04	Lab 3.0 °C 3.4 °C 4.4 °C 4.7 °C	Onsite 3.0 °C 3.4 °C 4.4 °C 4.7 °C
		Lab 3.0 °C 3.2 °C 4.4 °C	Onsite 3.0 °C 3.2 °C 4.4 °C
E8. (a) Temperature enclosures (Oven/Freezer/Incubator/Furnace) -80 °C to 100 °C 100 °C to 200 °C 200 °C to 400 °C 400 °C to 600 °C 600 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1100 °C (b) Temperature & Humidity enclosures (Chamber) 10 ~ 25 °C / up to 20 %r.h. 10 ~ 25 °C / 20 ~ 50 %r.h. 10 ~ 25 °C / 50 ~ 90 %r.h. 25 ~ 50 °C / up to 20 %r.h. 25 ~ 50 °C / 20 ~ 50 %r.h. 25 ~ 50 °C / 50 ~ 90 %r.h. (c) System Accuracy test (Temperature) -80 °C to 200 °C 200 °C to 400 °C 400 °C to 800 °C 800 °C to 1000 °C 1000 °C to 1100 °C System Accuracy test (Temperature/Humidity) (10 ~ 50) °C / (20 ~ 90) %r.h.	Opus in-house procedure OPCP-070 Issue 04 Opus in-house procedure OPCP-070 Issue 04 Opus in-house procedure OPCP-070 Issue 04	Lab 1.7 °C 2.1 °C 2.9 °C 3.3 °C 4.1 °C 5.1 °C 5.6 °C	Onsite 1.7 °C 2.1 °C 2.9 °C 3.3 °C 4.1 °C 5.1 °C 5.6 °C
		Lab 0.9 °C 0.9 °C 0.9 °C 1.0 °C 1.0 °C 1.0 °C	Onsite 3.1 %r.h. 4.2 %r.h. 6.3 %r.h. 3.1 %r.h. 4.1 %r.h. 5.4 %r.h.
Lab 1.5 °C 2.3 °C 3.5 °C 4.9 °C 5.5 °C	Onsite 1.5 °C 2.3 °C 3.5 °C 4.9 °C 5.5 °C		
Lab 0.5 °C	Onsite 3.0 %r.h.		

Schedule



Certificate No. : LA-2005-0324-C

Issue No. : 23

Date : 07 July 2025

Page : 24 of 24

MEASURED QUANTITIES / INSTRUMENTS/RANGE TO BE CALIBRATED	METHOD	CALIBRATION & MEASUREMENT CAPABILITY (CMC*)
E9. Non-contact thermometer Emissivity ($\epsilon=0.95$) Up to 50 °C 50 °C to 100 °C 100 °C to 200 °C 200 °C to 300 °C 300 °C to 400 °C 400 °C to 500 °C	Opus in-house procedure OPCP-071 Issue 04	Lab 0.9 °C 1.1 °C 2.3 °C 2.9 °C 3.6 °C 5.3 °C

* CMC is expressed as an expanded uncertainty estimated at a level of confidence of approximately 95%.

Approved Signatories

Mr Peter Foo - For all items except E2

Mr R Senthil - For Categories B, C, D, E

Note :

This laboratory is accredited in accordance with the recognized International Standard ISO/IEC 17025. A laboratory's fulfilment of the requirements of ISO/IEC 17025 means the laboratory meets both the technical competence requirements and **management system requirements** that are necessary for it to consistently deliver technically valid calibrations. The **management system requirements** in ISO/IEC 17025 are written in language relevant to laboratory operations and operate generally in accordance with the principles of ISO 9001.